

MEDICAL POLICY



MEDICAL POLICY DETAILS	
Medical Policy Title	MAMMOGRAPHY: COMPUTER- AIDED DETECTION (CAD)
Policy Number	6.01.23
Category	Technology Assessment
Effective Date	03/28/02
Revised Date	04/24/03, 05/19/04, 04/21/05, 02/16/06, 12/21/06
Archived Date	10/18/07
Edited Date	12/18/08, 11/19/09, 09/16/10, 09/15/11, 09/20/12, 09/19/13, 08/21/14, 09/17/15, 07/21/16, 07/20/17, 08/16/18, 08/15/19
Product Disclaimer	<ul style="list-style-type: none"> • If a product excludes coverage for a service, it is not covered, and medical policy criteria do not apply. • If a commercial product (including an Essential Plan product) or a Medicaid product covers a specific service, medical policy criteria apply to the benefit. • If a Medicare product covers a specific service, and there is no national or local Medicare coverage decision for the service, medical policy criteria apply to the benefit.

POLICY STATEMENT

Based upon our criteria and review of the peer-reviewed literature, digitalized computer-aided detection (CAD) mammography has been medically proven to be effective and therefore **medically appropriate** to identify and mark regions of interest on mammographic views to bring them to the attention of the radiologist after the initial reading has been completed.

Refer to Corporate Medical Policy #6.01.22 regarding Mammography: Digital Breast Tomosynthesis.

POLICY GUIDELINES

This policy does not address *digital mammography*. Digital mammography produces digital images as an option to conventional film screen mammography. Refer to Corporate Medical Policy 6.01.22, Mammography: Digital Breast Tomosynthesis.

DESCRIPTION

Computer-aided detection (CAD) has been suggested as an adjunct to screening mammograms to decrease errors in perception (e.g., failure to see abnormality). The use of CAD systems requires a digital image, either generated by digitization of a prior screen-film mammogram (digitized mammogram), or generated directly (direct full-field digital mammogram). Commercially available CAD systems then use computerized algorithms for identifying suspicious regions of interest on the digital image. The locations of the abnormalities are marked so the reader can then reference the same areas in the original mammogram for further review. The intent of CAD is to aid in detection of potential abnormalities for the radiologist to re-review. The radiologist, not CAD, makes the determination if a clinically significant abnormality exists and whether further diagnostic evaluation is warranted.

RATIONALE

Several CAD mammography systems have been approved by the U.S. Food and Drug Administration (FDA) for use in the U.S., including but not limited to the Cenova™ (Hologic, Inc.), MammoReader® (Intelligent Systems Software, Inc.), PowerLook Advanced Mammography Platform (iCAD Inc.), and RapidScreen™ RS-2000 (Deus Technologies).

The available evidence suggests that the use of CAD as an adjunct to the radiologists' interpretation of screen-film mammography improves net health outcomes compared with single reader radiologist interpretations, by increasing true positive rate without a disproportionate increase in the false positive rate.

Medical Policy: MAMMOGRAPHY: COMPUTER- AIDED DETECTION (CAD)

Policy Number: 6.01.23

Page: 2 of 4

A prospective study (Birdwell 2005) of 8,682 patients assessed the effect of CAD on screening mammogram interpretation by comparing radiologist interpretation of screen film mammograms followed by immediate re-evaluation of the mammograms with CAD information. Authors concluded that prospective clinical use of CAD resulted in a 7.4% increase in cancers detected.

Another prospective study of 6,111 women (Khoo 2005) evaluated recall and cancer detection rates with and without CAD. Screen film mammograms for 6,111 women were digitized and analyzed using CAD and then all mammograms were independently double-read by a consulting radiologist. Authors concluded that CAD increased sensitivity of single reading by 1.3%, however, double reading increased sensitivity by 8.2%.

A third prospective, historically controlled study (Cupples 2005) assessed the effect of CAD on screening outcomes (recall, biopsy and cancer detection rates in addition to type, size and stage of cancers found) by comparing 7,872 screening mammograms without CAD with 19,402 screening mammograms interpreted with CAD. Authors concluded that an increased detection rate, a younger age at diagnosis, and a significantly earlier stage of invasive cancer detection were consistent with a positive screening impact of CAD.

CODES

- *Eligibility for reimbursement is based upon the benefits set forth in the member’s subscriber contract.*
- ***CODES MAY NOT BE COVERED UNDER ALL CIRCUMSTANCES. PLEASE READ THE POLICY AND GUIDELINES STATEMENTS CAREFULLY.***
- *Codes may not be all inclusive as the AMA and CMS code updates may occur more frequently than policy updates.*

CPT Codes

Code	Description
77065	Diagnostic mammography, including computer-aided detection (CAD) when performed; unilateral
77066	Diagnostic mammography, including computer-aided detection (CAD) when performed; bilateral
77067	Screening mammography, bilateral (2-view study of each breast), including computer-aided detection (CAD) when performed

Copyright © 2019 American Medical Association, Chicago, IL

HCPCS Codes

Code	Description
No specific code(s)	

ICD10 Codes

Code	Description
C50.011- C50.019	Malignant neoplasm of nipple and areola, female (code range)
C50.111- C50.119	Malignant neoplasm of central portion of breast, female (code range)
C50.211- C50.219	Malignant neoplasm of upper-inner quadrant of breast, female (code range)
C50.311- C50.319	Malignant neoplasm of lower-inner quadrant of breast, female (code range)

Medical Policy: MAMMOGRAPHY: COMPUTER- AIDED DETECTION (CAD)**Policy Number: 6.01.23****Page: 3 of 4**

Code	Description
C50.411- C50.419	Malignant neoplasm of upper-outer quadrant of breast, female (code range)
C50.511- C50.519	Malignant neoplasm of lower-outer quadrant of breast (code range)
C50.611- C50.619	Malignant neoplasm of axillary tail of breast, female (code range)
C50.811- C50.819	Malignant neoplasm of overlapping site of breast (code range)
C50.911- C50.919	Malignant neoplasm of breast unspecified site, female (code range)
C79.81	Secondary malignant neoplasm of breast
C79.89	Secondary malignant neoplasm of other specified sites
C79.9	Secondary malignant neoplasm of unspecified site
D05.00-D05.92	Carcinoma in situ of breast (code range)
D48.60-D48.62	Neoplasm of uncertain behavior of breast (code range)
D49.3	Neoplasm of unspecified behavior of breast
R92.8	Other abnormal and inconclusive findings on diagnostic imaging of breast
Z12.31	Encounter for screening mammogram for malignant neoplasm of breast
Z12.39	Encounter for other screening for malignant neoplasm of breast
Z80.3	Family history of malignant neoplasm of breast
Z85.3	Personal history of malignant neoplasm of breast

REFERENCES

Birdwell RL, et al. Computer-aided detection with screening mammography in a university hospital setting. Radiol 2005 Aug;236(2):451-7.

BlueCross BlueShield Association. Computer-aided detection in conjunction with digitized screen-film mammography or full-field digital mammography. Medical Policy Reference Manual Policy # 6.01.39. Archived 2009 Dec 10.

*BlueCross BlueShield Association Technology Evaluation Center (TEC). Computer-aided detection in mammography. Technology Assessment Program. 2002 Dec;17(17).

*BlueCross BlueShield Association Technology Evaluation Center (TEC). Computer-aided detection with full-field digital mammography. Technology Assessment Program. 2006 May;21(3).

Brem RF, et al. Impact of breast density on computer-aided detection for breast cancer. AJR Am J Roentgenol 2005 Feb;184(2):439-44.

*Brem RF, et al. Improvement in sensitivity of screening mammography with computer-aided detection: a multi-institutional trial. AJR Am J Roentgenol 2003 Sep;181(3):687-93.

Cawson JN, et al. Invasive breast cancers detected by screening mammography: a detailed comparison of computer-aided detection-assisted single reading and double reading. J Med Imaging Radiat Oncol 2009 Oct;53(5):442-9.

Dean JC, et al. Improved cancer detection using computer-aided detection with diagnostic and screening mammography: prospective study of 104 cancers. AJR Am J Roentgenol 2006 Jul;187(1):20-8.

Medical Policy: MAMMOGRAPHY: COMPUTER- AIDED DETECTION (CAD)

Policy Number: 6.01.23

Page: 4 of 4

Destounis SV, et al. Can computer-aided detection with double reading of screening mammograms help decrease the false-negative rate? Initial experience. Radiol 2004 Aug;232(2):578-84.

Fenton JJ, et al. Effectiveness of computer-aided detection in community mammography practice. J Natl Cancer Inst. 2011 Aug 3;103(15):1152-1161.

Fenton JJ, et al Influence of computer-aided detection on performance of screening mammography. NEJM 2007 Apr 5;356(14):1399-409.

Gilbert FJ, et al. single reading with computer-aided detection for screening mammography. N Eng J Med 2008 Oct 16;359(16):1675-84.

Gromet Matthew. Comparison of computer-aided detection to double reading of screening mammograms: review of 231,221 mammograms. AJR Am J Roentgenol 2008 Apr;190(4):854-9.

Kim SJ, et al. Computer-aided detection in full-field digital mammography: sensitivity and reproducibility in serial examinations. Radiology 2008 Jan;246(1):71-80.

Noble M., et al. Computer-aided detection mammography for breast cancer screening: systematic review and meta-analysis. Arch Gynecol Obstet 2009 Jun;279(6):881-90.

*Key Article

KEY WORDS

CAD mammography, Digitalized computer-aided detection mammography.

CMS COVERAGE FOR MEDICARE PRODUCT MEMBERS

Based on our review, computer-aided detection mammography is not addressed in National or Regional Medicare coverage determinations or policies.