

MEDICAL POLICY



MEDICAL POLICY DETAILS	
Medical Policy Title	VENTRICULAR ASSIST DEVICES
Policy Number	7.01.07
Category	Technology Assessment
Effective Date	10/18/01
Revised Date	09/19/02, 09/16/04, 07/21/05, 07/20/06, 05/17/07, 05/14/08, 05/28/09, 05/27/10, 05/19/11, 05/24/12, 06/20/13, 05/22/14, 06/18/15, 06/16/16, 06/15/17, 06/21/18, 08/14/19, 08/15/19
Product Disclaimer	<ul style="list-style-type: none"> • If a product excludes coverage for a service, it is not covered, and medical policy criteria do not apply. • If a commercial product (including an Essential Plan product) or a Medicaid product covers a specific service, medical policy criteria apply to the benefit. • If a Medicare product covers a specific service, and there is no national or local Medicare coverage decision for the service, medical policy criteria apply to the benefit.

POLICY STATEMENT

Based upon our criteria and review of the peer-reviewed literature, FDA approved ventricular assist devices have been medically proven to be effective and therefore, **medically appropriate** for the following indications:

- I. When used as a bridge to transplantation for patients diagnosed with severe ventricular heart failure who meet all of the following criteria:
 - A. Approved as heart transplant candidate by an approved heart transplant center; and
 - B. An imminent risk of dying before donor heart procurement; and
 - C. On optimal inotropic (influencing the contractility of muscular tissue) support; and,
 - D. On an intra-aortic balloon pump (IABP), unless contraindicated.
- II. As a bridge to recovery for post-cardiotomy patients who are unable to be weaned from cardiopulmonary bypass, or for patients with potentially reversible left ventricular dysfunction due to acute cardiogenic shock or acute myocarditis.
- III. As destination therapy for patients with chronic end-stage cardiac failure who meet all of the following criteria based on the REMATCH study:
 - A. Evaluated as a heart transplant candidate by an approved heart transplant center; but determined to be ineligible for heart transplantation; and
 - B. NYHA class III or IV heart failure for at least 28 days who have received at least 14 days support with an intraaortic balloon pump or are dependent on IV inotropic agents OR class IV end-stage heart failure that has failed to respond to optimal medical management for at least 60 of the last 90 days; and
 - C. Functional limitation with a peak oxygen consumption of less than or equal to 14 ml/kg/min; and
 - D. Appropriate body size to support the LVAD implantation.
- IV. Percutaneous ventricular assist devices are considered **investigational** for all indications.

Refer to Corporate Medical Policy #7.01.65 regarding Artificial Hearts.

POLICY GUIDELINES

- I. The following guidelines may be used as hemodynamic selection criterion for bridge to transplant:
 - A. Either a left atrial pressure of 20m Hg or a cardiac index of less than 2.0L/min/m;
 - B. Patients are generally being treated as inpatients and categorized by the American Heart Association, or comparable, as Class IV CHF; and
 - C. Classified as Status I by the United Network for Organ Sharing (considered the highest priority for transplantation).

Medical Policy: VENTRICULAR ASSIST DEVICES

Policy Number: 7.01.07

Page: 2 of 8

- II. Contraindications for bridge to transplant:
 - A. Includes conditions which would generally exclude patients for heart transplant:
 - 1. Chronic irreversible hepatic, renal, or respiratory failure, or
 - 2. Systemic infection, or
 - 3. Blood dyscrasia.
 - B. Uncorrected valvular diseases, due to the potential problems with adequate function of the ventricular assist device.
- III. Individuals considered for VAD implantation, as destination therapy should be evaluated for potential difficulties that would complicate and diminish the success of the implantation including an assessment of patient compliance.
- IV. The DeBakey VAD Child (HeartAssist[®] 5 pediatric VAD) has Humanitarian Device Exemption authorization for use in providing temporary left side mechanical circulatory support as a bridge to cardiac transplantation for pediatric patients (5-16 years old, with BSA greater than or equal 0.7 m² and less than 1.5 m²) who are in NYHA Class IV end stage heart failure, are refractory to medical therapy and who are (listed) candidates for cardiac transplantation. The Berlin Heart EXCOR VAD is indicated for children with severe isolated left ventricular or biventricular dysfunction who are candidates for cardiac transplant and require circulatory support and has also received FDA approval through the humanitarian device exemption (HDE) process.

DESCRIPTION

Ventricular assist devices (VADs) fit into the general category of mechanical circulatory assist devices. VADs have been developed to provide mechanical support for patients with severe heart failure who are awaiting a heart transplant (bridge to transplant), for patients with post cardiectomy or potentially reversible left ventricular dysfunction (bridge to recovery), and in certain specific instances, for patients with end stage heart failure who are not suitable transplant candidates (destination therapy).

Bridging to heart transplantation involves improving hemodynamics and restoring organ function such that a patient may have a better probability of surviving until a donor heart is available. Destination therapy is used for individuals with end stage heart failure (who are not candidates for heart transplant), who are currently receiving optimal medical therapy with ACE inhibitors, beta-blockers, and inotropic drugs. Left ventricular assist devices (LVADs) are also used temporarily for post-cardiectomy patients who cannot be weaned from cardiopulmonary bypass immediately following surgery. VADs have also been investigated as a bridge to recovery in patients with potentially reversible left ventricular dysfunction due to acute cardiogenic shock or acute myocarditis.

RATIONALE

VADs can provide an effective bridge to transplantation, allowing patients to survive until a donor heart is available. Published studies report that use of a VAD does not compromise the success of subsequent heart transplantation, and may actually improve post-transplant survival.

The use of VADs as destination therapy is supported by the REMATCH study, a randomized controlled trial that compared LVAD device transplantation with optimal medical management in 129 patients with end-stage heart failure who were not candidates for cardiac transplantation. The trial showed that patients who received a VAD had a longer survival rate than those treated with optimal medical therapy. Median survival was increased by approximately 8.5 months. Although adverse events were more likely in the VAD group, these appeared to be outweighed by better outcomes on function; NYHA class was significantly improved, as was quality of life among those living to 12 months. Two years of additional observation on REMATCH patients (Park, 2005) substantiates the continuing survival benefit of LVAD support. LVAD treatment more than doubled the survival seen at 2 years over optimal medical management.

VADs have been used as bridge to recovery for patients with potentially reversible left ventricular dysfunction. Implantation of VADs provides circulatory support and allows myocardial recovery in post cardiectomy cases where the patient cannot be weaned from cardiopulmonary bypass, and in patients with acute cardiogenic shock or acute myocarditis.

Medical Policy: VENTRICULAR ASSIST DEVICES

Policy Number: 7.01.07

Page: 3 of 8

A variety of devices have received approval for marketing from the U.S. Food and Drug Administration (FDA), encompassing both biventricular and left ventricular devices. The type of device used is dependent upon specific FDA-labeled indications. These devices include, but are not limited to:

- I. HeartMate® Sutures Not APplied Vented Electric Left Ventricular Assist System (SNAP VE LVAS) (Thoratec Corp.);
- II. HeartMate® II LVAD (Thoratec Corp.);
- III. HeartMate 3™ Left Ventricular Assist System (Thoratec Corp.);
- IV. Impella 2.5®/ 5.0® (AbioMed Cardiovascular, Inc);
- V. Abiomed BVS® 5000 Biventricular Support System (AbioMed Cardiovascular, Inc.).

Two pulsatile devices, the HeartMate SNAP VE LVAS, and the HeartMate XVE LVAS, have received FDA approval as destination therapy. The Heartmate II LVAD, a continuous flow device received FDA approval as destination therapy January 20, 2010. The premarket approval included two-year data from a study cohort of 200 patients randomly assigned 2:1 to either a HeartMate XVE or a HeartMate II. Patients implanted with the HeartMate II device had statistically significant improved two-year survival vs. those implanted with the HeartMate XVE, in addition to improved quality of life. Forty-six percent of the 134 patients implanted with the HeartMate II were still living after two years with no disabling stroke or need for reoperation, device replacement or repair, compared with 11% in the 66 patient control group. Approval was contingent on a post-approval follow-up study involving 247 patients for either two-years or until outcome by Thoratec.

In February 2004, the FDA approved a Humanitarian Device Exemption (HDE) for the DeBakey VAD® Child, a ventricular assist device for home and hospital use, for children aged 5 to 16 awaiting a heart transplant. The FDA approved the HDE for the VAD after reviewing data that showed the device had a reasonable probability of being safe and effective in children. Publications have reported positive outcomes for children using VADs as a bridge to transplantation. Using the UNOS database, Davies, et al. (2008) reported on use of VADs in pediatric patients undergoing heart transplantation. Their analysis concluded that pediatric patients requiring a pretransplantation VAD have similar long-term survival to those not receiving mechanical circulatory support.

Prior to April 2008, only pulsatile LVADS devices were FDA approved for long-term use. Non-pulsatile axial flow devices are smaller in size and have other technical advantages over pulsatile models. The HeartMate II (Thoratec) is the first continuous flow device to receive FDA approval as a bridge to transplant (BTT) for treatment of advanced-stage heart failure. The approval was based on one-year follow-up data from the first 194 HeartMate II BTT patients enrolled in the trial. Results included in the final PMA submission were:

- I. The median duration of support was 132 days, and the cumulative patient support in the trial was 109 years.
- II. Survival to cardiac transplantation, recovery or ongoing on HeartMate II support was 80 % at six months and 77 % at one year.
- III. 84 % of the patients survived to hospital discharge or transplantation.
- IV. Significant improvements were observed across all measures of functional status and quality of life as compared to baseline status.
- V. The incidence of major adverse events with comparable definitions - including infections, strokes and bleeding requiring surgery - was significantly lower than what was clinically observed in the previous BTT study of the HeartMate VE LVAS.

The HeartMate 3™ Left Ventricular Assist System (Thoratec) was approved by the FDA on August 23, 2017. Per the manufacturer's website, the HeartMate 3 system can pump up to 10 liters of blood per minute and is the only commercially-approved continuous flow implantable left ventricular assist system to utilize Full MagLev™ (fully magnetically-levitated) Flow technology, which allows the device's rotor to be "suspended" by magnetic forces—rather than bearings—with the goal of being able to more gently pass the blood cells through the pump. The magnets keep the rotor in place by calibrating tens of thousands of times per second to ensure it stays suspended and centered within the pump, no matter the speed settings used by a physician. This ensures the pump is performing effectively while continuing to deliver the best patient therapy possible. The HeartMate 3 system also uses the industry's widest pump pathway, designed so the blood cells are not damaged when passing through. The system also relies on a built-in "pulse" programmed to help ensure the blood continues to move through without becoming static, therefore reducing the risk of

Medical Policy: VENTRICULAR ASSIST DEVICES

Policy Number: 7.01.07

Page: 4 of 8

blood clot formation. The HeartMate 3 blood pump should not be used in patients who cannot tolerate, or who are allergic to, anticoagulation therapy (blood thinners) because these medicines are required to prevent blood clots from forming in the pump.

The Jarvik 2000, a non-pulsatile axial flow blood pump, is in phase II and III clinical trials. The Jarvik 2000 is used by hospitals in the United States as a bridge to heart transplant under an FDA-approved clinical investigation. In Europe, the Jarvik 2000 has earned CE Mark certification for both bridge-to-transplant and lifetime use. As an investigational device, the Jarvik 2000 has been implanted in more than 200 patients dying of heart failure.

AbioMed, Inc announced the FDA 510(k) clearance for its Impella 2.5 cardiac assist device June 2, 2008. The Impella 2.5 is inserted percutaneously via the femoral artery into the left ventricle to provide partial circulatory support for periods up to six hours. Up to 2.4 liter of blood per minute is delivered by the pump from the left ventricle into the ascending aorta, providing the heart with active support in critical situations. The PROTECT I trial (Dixon, et al. 2009) evaluated the effectiveness of the Impella 2.5 (n = 20) in patients undergoing high risk PCI at seven centers. Eligible patients had a LEF of less than 35%. The Impella 2.5 device was implanted successfully in all patients. The mean duration of circulatory support was 1.7 ± 0.6 h (range: 0.4 to 2.5 h). Mean pump flow during PCI was 2.2 ± 0.3 l/min. At 30 days, the incidence of major adverse cardiac events was 20% (2 patients had a periprocedural myocardial infarction; 2 patients died at days 12 and 14). There was no evidence of aortic valve injury, cardiac perforation, or limb ischemia. Two patients (10%) developed mild, transient hemolysis without clinical sequelae. None of the patients developed hemodynamic compromise during PCI. Other studies investigating the Impella device, although limited by small sample populations, have demonstrated its efficacy in providing circulatory support during high-risk percutaneous revascularization procedures and in post-cardiotomy patients.

The Society for Heart and Lung Transplantation (Gronda, 2006), published guidelines for the care of cardiac transplant candidates that included considerations for the use of ventricular assist devices. The recommendations for VAD therapy are based on a comparison of short- and long-term survival and QOL outcomes with conventional therapy and are available at: [http://www.jhltonline.org/article/S1053-2498\(06\)00251-8/pdf](http://www.jhltonline.org/article/S1053-2498(06)00251-8/pdf).

CODES

- *Eligibility for reimbursement is based upon the benefits set forth in the member's subscriber contract.*
- **CODES MAY NOT BE COVERED UNDER ALL CIRCUMSTANCES. PLEASE READ THE POLICY AND GUIDELINES STATEMENTS CAREFULLY.**
- *Codes may not be all inclusive as the AMA and CMS code updates may occur more frequently than policy updates.*

CPT Codes

Code	Description
33975	Insertion of ventricular assist device; extracorporeal, single ventricle
33976	Insertion of ventricular assist device; extracorporeal, biventricular
33977	Removal of ventricular assist device; extracorporeal, single ventricle
33978	Removal of ventricular assist device; extracorporeal, biventricular
33979	Insertion of ventricular assist device, implantable intracorporeal, single ventricle
33980	Removal of ventricular assist device, implantable intracorporeal, single ventricle
33981	Replacement of extracorporeal ventricular assist device, single or biventricular, pump(s), single or each pump
33982	Replacement of ventricular assist device pump(s) implantable, intracorporeal, single ventricle, without cardiopulmonary bypass
33983	Replacement of ventricular assist device pump(s); implantable intracorporeal, single ventricle, with cardiopulmonary bypass
33990 (E/I)	Insertion of ventricular assist device, percutaneous including radiological supervision and interpretation; arterial access only
33991 (E/I)	Insertion of ventricular assist device, percutaneous including radiological supervision and interpretation; both arterial and venous access, with transeptal puncture

Medical Policy: VENTRICULAR ASSIST DEVICES**Policy Number: 7.01.07****Page: 5 of 8**

Code	Description
33992 (E/I)	Removal of percutaneous ventricular assist device at separate and distinct session from insertion
93750	Interrogation of ventricular assist device (VAD), in person, with physician or other qualified health care professional analysis of device parameters (eg, drivelines, alarms, power surges), review of device function (eg, flow and volume status, septum status, recovery), with programming, if performed, and report

*Copyright © 2019 American Medical Association, Chicago, IL***HCPCS Codes**

Code	Description
Q0477	Power module patient cable for use with electric or electric/pneumatic ventricular assist device, replacement only
Q0480-Q0509	VAD components (code range)

ICD10 Codes

Code	Description
A18.84	Tuberculosis of heart
I09.81	Rheumatic heart failure
I11.0	Hypertensive heart disease with heart failure
I13.0	Hypertensive heart and chronic kidney disease with heart failure and stage 1 through stage 4 chronic kidney disease, or unspecified chronic kidney disease
I13.2	Hypertensive heart and chronic kidney disease with heart failure and with stage 5 chronic kidney disease, or end stage renal disease
I21.01-I22.9	ST elevation (STEMI) and non-ST elevation (NSTEMI) myocardial infarction (code range)
I40.0-I41	Acute myocarditis and Myocarditis in diseases classified elsewhere (code range)
I50.1-I50.9	Heart failure (code range)
R57.0	Cardiogenic shock

REFERENCES

Acharya D, et al. Ventricular assist device in acute myocardial infarction. J Am Coll Cardiol 2016 Apr 26;67(16):1871-80.

Adamo L, et al. The Heartmate risk score predicts morbidity and mortality in unselected left ventricular assist device recipients and risk stratifies INTERMACS Class 1 patients. JACC Heart Fail 2015 Apr;3(4):283-90.

Almond CS, et al. Berlin heart EXCOR pediatric ventricular assist device for bridge to heart transplantation in US children. Circulation 2013 Apr 23;127(16):1702-11.

BlueCross BlueShield Association. Ventricular assist devices and total artificial hearts. Medical Policy Reference Manual Policy # 7.03.11 2018 Aug 09.

*BlueCross BlueShield Association Technology Evaluation Center (TEC). Left ventricular devices as destination therapy for end-stage heart failure. 2002 Dec;17(19).

*BlueCross Blue Shield Association Technology Assessment (TEC) Special report: cost effectiveness of left-ventricular assist devices as destination therapy for end-stage heart failure. 2004 Apr; 19(2).

*Blume ED, et al. Outcomes of children bridged to heart transplantation with ventricular assist devices: a multi-institutional study. Circ 2006 May 16;113(19):2313-9.

*Brancaccio G, et al. Ventricular assist devices as a bridge to heart transplantation or as destination therapy in pediatric patients. Transplant Proc 2012 Sep;44(7):2007-12.

Medical Policy: VENTRICULAR ASSIST DEVICES

Policy Number: 7.01.07

Page: 6 of 8

- *Clegg AJ, et al. Clinical and cost-effectiveness of left ventricular assist devices as destination therapy for people with end-stage heart failure: a systematic review and economic evaluation. Int J Technol Assess Health Care 2007 Spring;23(2):261-8.
- *de By TM, et al. The European Registry for Patients with Mechanical Circulatory Support (EUROMACS): first annual report. Eur J Cardiothorac Surg 2014 May;47(5):770-7.
- *Donneyong M, et al. The association of pre-transplant HeartMate® II left ventricular assist device implantation. Circulation 2014 Mar 11;129(10):1161-6.
- *Dunlay SM, et al. Frailty and outcomes after implantation of left ventricular assist device as destination therapy. J Heart Lung Transplant 2014 Apr;33(4):359-65.
- Fendler TJ, et al. Incidence and predictors of cognitive decline in patients with left ventricular assist devices. Circ Cardiovasc Qual Outcomes 2015 May; 8(3): 285–291.
- *Fraser CD, et al. Prospective trial of a pediatric ventricular assist device. N Engl J Med 2012 Aug 9;367(6):532-41.
- Grimm JC, et al. Outcomes in patients bridged with univentricular and biventricular devices in the modern era of heart transplantation. Ann Thorac Surg 2016 Jul;102(1):102-8.
- *Gronda E, et al. Heart rhythm considerations in heart transplant candidates and considerations for ventricular assist devices: International Society for Heart and Lung Transplantation guidelines for the care of cardiac transplant candidates--2006. J Heart Lung Transplant 2006 Sep;25(9):1043-56.
- Hanke JS, et al. One-year outcomes with the HeartMate 3 left ventricular assist device. J Thorac Cardiovasc Surg 2018 Aug;156(2):662-669.
- *Hetzer R, et al. Role of paediatric assist device in bridge to transplant. Ann cardiothorac Surg 2018;7(1):82-98.
- *Hetzer R, et al. Single-center experience with treatment of cardiogenic shock in children by pediatric ventricular assist devices. J Thorac Cardiovasc Surg 2011 Mar;141(3):616-623. e1.
- Hu FB and Cui LQ. Percutaneous left ventricular assist device vs. intra-aortic balloon pump in patients with severe left ventricular dysfunction undergoing cardiovascular intervention: a meta-analysis. Chronic Dis Transl Med 2018 Apr 12;4(4):260-267.
- Jakovljevic DG, et al. Left ventricular assist device as a bridge to recovery for patients with advanced heart failure. JACC 2017;69(15):1924-33.
- *Jorde UP, et al. Results of the destination therapy post-FDA-approval study with a continuous flow left ventricular assist device: a prospective study using the INTERMACS Registry. J Am Coll Cardio 2014 May 6;63(17):1751-7.
- *Lietz K, et al. Outcomes of left ventricular assist device implantation as destination therapy in the post-REMATCH era: implications for patient selection. Circulation 2007 Jul 31;116(5):497-505.
- Lima B, et al. Controversies and challenges of ventricular assist device therapy. Am J Cardiol. 2018;121(10):1219-1224.
- Loehn T, et al. Long term survival after early unloading with Impella CP in acute myocardial infarction complicated by cardiogenic shock. Eur Heart J Acute Cardiovasc Care 2018 Nov20 [Epub ahead of print].
- Luni FK, et al. Percutaneous left ventricular assist device support during ablation of ventricular tachycardia: a meta-analysis of current evidence. J Cardiovasc Electrophysiol 2019 Mar 7 [Epub ahead of print].
- Mehra MR, et al. Two-year outcomes with a magnetically levitated cardiac pump in heart failure. NEJM 2018;378(15):1386-95.
- Miller KW and Rogers JG. Evolution of left ventricular assist device therapy for advanced heart failure: a review. JAMA Cardiol 2018;3(7):650-658.
- *Morales DL, et al. Bridging children of all sizes to cardiac transplantation: the initial multicenter North American experience with the Berlin Heart EXCOR ventricular assist device. J Heart Lung Transplant 2011 Jan;30(1):1-8.

Medical Policy: VENTRICULAR ASSIST DEVICES

Policy Number: 7.01.07

Page: 7 of 8

*Moreno SG, et al. Cost-effectiveness of the implantable HeartMate II left ventricular assist device for patients awaiting heart transplantation. J Heart Lung Transplant 2012 May;31(5):450-8.

Morgan JA, et al. Outcomes on continuous flow left ventricular assist devices: a single institutional 9-year experience. Ann Thorac Surg 2016;102:1266–73.

*Morgan JA, et al. Stroke while on long-term left ventricular assist device support: incidence, outcome, and predictors. ASAIO J 2014 May-Jun;60(3):284-9.

O'Neill WW, et al. analysis of outcomes for 15,259 US patients with acute myocardial infarction cardiogenic shock (AMICS) supported with the Impella device. Am Heart J 2018 Aug;202:33-38.

Ouweneel DM, et al. Real-life use of left ventricular circulatory support with Impella in cardiogenic shock after acute myocardial infarction: 12 years AMC experience. Eur Heart J Acute Cardiovasc Care 2019 Jun;8(4):338-349.

*Patel CB, et al. A contemporary review of mechanical circulatory support. J Heart Lung Transplant 2014 Jul;33(7):667-74.

Patel SR, et al. Heart transplantation versus continuous-flow left ventricular assist device: comprehensive cost at 1 year. J Card Fail 2015 Apr;3(4):282-90.

*Rogers JG, et al. Chronic mechanical circulatory support for inotrope-dependent heart failure patients who are not transplant candidates: results of the INTrEPID Trial. J Am Coll Cardiol 2007 Aug 21;50(8):741-7.

Rogers JG, et al. Intrapericardial left ventricular assist device for advanced heart failure. NEJM 2017;376(5):451-60.

Seco M, et al. Long-term prognosis and cost-effectiveness of left ventricular assist device as bridge to transplantation: a systematic review. Int J Cardiol 2017 May 15;235:22-32.

*Sharma MS, et al. Ventricular assist device support in children and adolescents as a bridge to heart transplantation. Ann Thorac Surg 2006 Sep;82(3):926-32.

*Shuhaiber JH, et al. The influence of preoperative use of ventricular assist devices on survival after heart transplantation: propensity score matched analysis. BMJ 2010 Feb10;340:c392.

*Starling RC, et al. results of the post-U.S. Food and Drug Administration-approval study with a continuous flow left ventricular assist device as a bridge to heart transplantation: a prospective study using the INTERMACS (Interagency Registry for Mechanically Assisted Circulatory Support). J Am Col Cardiol 2011 May 10;57(19):1890-8.

Takeda K, et al. Outcome of cardiac transplantation in patients requiring prolonged continuous-flow left ventricular assist device support. J Heart Lung Transplant 2015 Jan;34(1):89-99.

Theochari CA, et al. Heart transplantation versus left ventricular assist devices as destination therapy or bridge to transplantation for 1-year mortality: a systematic review and meta-analysis. Ann cardiothorac Surg 2018;7(1):3-11.

Yin MY, et al. Post-transplant outcome in patients bridged to transplant with temporary mechanical circulatory support devices. J Heart Lung Transplant 2019 Apr 20 [Epub ahead of print].

*Key Article

KEY WORDS

Bridge to heart transplant, Assist Devices, ventricular, LVAD, VAD, Destination Therapy

CMS COVERAGE FOR MEDICARE PRODUCT MEMBERS

There is currently a National Coverage Determination (NCD) for Artificial Hearts and Related Devices. Please refer to the following NCD website for Medicare Members: <https://www.cms.gov/medicare-coverage-database/details/ncd-details.aspx?NCDId=246&ncdver=6&bc=AgAAgAAAAAAAAA%3d%3d&>.

Medical Policy: VENTRICULAR ASSIST DEVICES

Policy Number: 7.01.07

Page: 8 of 8